

OCT 31 2005

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130492</u>	
Name <u>John T. Ostoj</u>		ID# <u>4823</u>	Date <u>10/26/05</u>
A	Agency <u>A ST</u>		Phone # _____
Instrument Location <u>Willow</u>			
B	Alco S/N <u>Y 301573</u> Target Value <u>083</u> High Pressure <u>800</u>		
Alco Test Value Average <u>083</u> <u>082</u> <u>11/16/05</u>			
1 st Alco 2 nd Alco			
Signature <u>[Signature]</u>			
(OVER)			

Under the Alaska Rules of Evidence, I certify that:
(Do Not write in the area below)

(a) I am Chris W. Beheim, Director, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Ostoj, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Chris W. Beheim 11-15-05
Chris W. Beheim Date
Scientific Director
State of Alaska
State Breath Alcohol Testing Program

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130492

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

OCTOBER 26, 2005

OPERATOR'S NAME:

OSTOJ/JOHN/T

OPERATOR'S NUMBER: 4823

SUBJECT'S LAST NAME:

VOC TEST

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: TEST

DEPT/AGENCY: TALO

CASE/REPORT: TEST

TEST TYPE: V

ALCO TARGET VALUE: .083

ALCO S/N: K301573

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 29.67 in

ALCO TARGET .082 19:58

BLANK TEST .000 19:59

INTERNAL STANDARD VERIFIED 19:59

ALCO TV 29.67 in .083 20:00

BLANK TEST .000 20:01

SUBJECT SAMPLE .000 20:01

BLANK TEST .000 20:02

ALCO TV 29.68 in .082 20:02

BLANK TEST .000 20:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

OCTOBER 26, 2005

TIME 20:04

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.67 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvmxyz{|}~